



How to Fight COVID-19 with Portable Ventilator High-flow nasal catheter oxygen therapy or non-invasive mechanical ventilation:

When the patients have the problem of respiratory distress and / or hypoxemia can't be relieved after receiving standard oxygen therapy, then the high-flow nasal catheter oxygen therapy or non-invasive ventilation can be considered. If the condition can't be improved or worsened within the short time (1-2 hours), the tracheal intubation and invasive mechanical ventilation should be performed in time.

## Original part



## Duo-level non-invasive Ventilator

## Portable Ventilator Expert for COVID-19

## Specification

Ventilation mode	CPAP, S. T. S/T PC AVAPS
Frequence	Adjusting range: 3/min-40/min
	Increment: 1/min
Inspiration time	Adjusting range: 0-3s
	Increment: 0.1s
IPAP	Adjusting range: 4cmH₂O-25cmH₂O
	Increment: 0.5cmH₂O
EPAP	Adjusting range: 4cmH <sub>2</sub> O-20cmH <sub>2</sub> O
	Increment: 0.5cmH₂O
CPAP	Adjusting range: 4cmH <sub>2</sub> O-20cmH <sub>2</sub> O
	Increment: 0.5cmH₂O
Ramp time	Adjusting range: 0-60min
	Increment: 1min
Rise time	Adjusting range: 0.1s-0.5s
	Increment: 0.1s
Inspiration trigger	Adjusting range: auto adjusting, 1-3
	Increment: 1
Expiration trigger	Adjusting range: auto adjusting, 1-3
	Increment: 1
Flex	Adjusting range: OFF, 1-3
	Increment: 1
Humidifier	Adjusting range: 0-5
	Increment: 1
Ppeak	0-30cmH₂O

Oscillogram	Airway pressure histogram
Alarm and protection	
Low ventilation	0-15L/min
High airway pressure	20cmH2O-30cmH₂O
High frequency	1/min-99/min
Low frequency	1/min-60/min
Suffocation warning	8s-12s no spontaneous ventilation
The maximum limited pressure	<12.5 kPa
Pipe detached	Pipe detached 16s
Working condition	
Environmental temperature	5℃-35℃
Relative humidity	15%-95%
Atmosphere pressure	830hPa-1060hPa
Using condition	
Storage conditions	-20°C-+60°C
Relative humidity	15%-95%
Atmosphere pressure	500hPa-1060hPa
Packing size	= = = = = = = = = = = = = = = = = = = =
Carton packing size	L 285*W 350*H 290mm
G.W.	4.5KG
N.W.	3.5KG





